

**Equipment safety checks ~ beginning of every shift**

**Waters circuit** attached to  $O_2$  +  
Cushioned mask  
**Bag valve mask** available

**Guedel airways &  
Yankauer suckers**

**Suction**  
Working and attached

**Tracheostomy**  
2 spares ~1 same size & 1 a half size smaller  
Tracheal dilators

**Monitor**  
Current parameters, alarm limits, volume

**Ventilator**  
Current parameters/alarm limits/ mode/ settings

**Tranducers**  
level, zero, pressure bag at 300mmHg, fluid

**IVs**  
Drug, concentration, rate

If in doubt call for help