






EXTERNAL PACING

- 1 Attach defibrillation pads and 3 lead monitoring
- 2 Press 
- 3 Press 
- 4 Press  to increase current until electrical capture
- 5 **Palpate central pulse to confirm mechanical capture**
- 6 If necessary, increase  until mechanical capture
- 7 To view intrinsic rhythm press **and hold** 
- 8 Consider sedation or analgesia if patient uncomfortable